



Oxygen session

8th BGC-Argo, 20th ADMT, Villefranche-sur-Mer, France

Henry Bittig (IOW)

Ken Johnson, Josh Plant, Tanya Maurer (MBARI)



Virginie Thierry (Ifremer), Catherine Schmechtig (CNRS)

et al.



Good news first: We already came quite some way! 😊

- 180k profiles with DOXY, of which 80k have received DMQC
- Well documented processing and quality control
- Approaching Winkler Quality in the best cases !



Processing: State of the art

- BGC-Argo DAC cookbook:
stable; tiny, ca. yearly amendments
- BGC-Argo QC manual:
essentials present (simple RTQC tests, data adjustment), future refinements expected (finer QC tests, adjusted error estimation, ...)

argo data management


Argo data management
<http://dx.doi.org/10.13155/39795>

Processing Argo OXYGEN data at the DAC level
Version 2.3.1
June 13th 2018

argo data management

Argo data management
DOI: <http://dx.doi.org/10.13155/46542>

Argo Quality Control Manual for Dissolved Oxygen Concentration
Version 2.0
23th October 2018




Processing: State of the art

- DAC: Processing raw data -> DOXY
- DAC: RTQC DOXY -> DOXY_ADJUSTED (after PI DMQC feedback)
- PI: Quick first DMQC for RT adjustment
- PI: Repeat DMQC in regular intervals

| | |
|----------------------|---|
| argo data management | Argo data management http://dx.doi.org/10.13155/39795 |
| | Processing Argo OXYGEN data at the DAC level Version 2.3.1 June 13th 2018 |

| | |
|----------------------|--|
| argo data management | Argo data management DOI: http://dx.doi.org/10.13155/46542 |
| | Argo Quality Control Manual for Dissolved Oxygen Concentration Version 2.0 23 th October 2018 |



Bittig et al. (2019) A BGC-Argo guide: Planning, deployment, data handling and usage. Front. Mar. Sci. 6:502.
doi: 10.3389/fmars.2019.00502

Agenda of refinements

Processing

- PhaseCoef0 issue

RTQC

- Quick biofouling with rapid daily cycling during day
- Need faster ways to filter out (probably) bad data quickly:
Regional Range Tests, Surface Saturation Test

Adjustments

- Adjustment maths, adjusted error estimate
- Workflow
- Time response correction (DM), MTIME variable

Follow-up: PhaseCoef0 issue

Following modification of cal. procedures after Apr. 2017, not all calibration coefficients were stated on Aanderaa 4330 calibration sheets. Fixed ~Sept. 2018.

Optode 4330 SN range ~2400 - ~3000

-> AO 103, IF 88, IN 10, HZ 7, JA 1 float

New deployments with 4330 optodes in this SN range need checking!

Betreff: [argo-bio] Processing Argo O2 Data and Argo O2 QC

Von: Henry Bittig <bittig@obs-vlfr.fr>

Datum: 14.06.2018 15:05

An: Argo Bio <argo-bio@jcommops.org>, argo-dm <argo-dm@jcommops.org>

Dear Argo-bio and Argo-dm lists,

please find attached the updated O2 cookbook and QC manual with modifications according to the last ADMT meeting and with the request for comments within the next three weeks (until July 5), after which they will be posted on the ADMT website.

Modifications to the versions distributed before ADMT18 in Hamburg are
O2 cookbook:

- There was a typo in the O2 solubility's B2 factor: Its value is $-1.0341e-2$, not $-1.0341e-3$.

The impact is minor and probably doesn't merit a full DOXY reprocessing in its own, but should be corrected for new data (all DACs).

- Aanderaa 4330 optode multi-point calibrations after Apr. 2017 may include a PhaseCoef0 different from 0. This is included in the optode internal config, but it is not stated on the calibration certificates. If the calibration certificates are used by a DAC, PhaseCoef0=0 or PhaseCoef0~0 must be verified by re-calculation of the calibration data. The value of PhaseCoef0 is retrievable from the calibration certificates (I can share Matlab code if needed).

- PARAMETER_ACCURACY gives realistic values for different optode calibrations. *They state the "plug & play" accuracy, not the accuracy attainable after adjustment.*

Follow-up: PhaseCoef0 issue – pre

Following modification of cal. procedures after Apr. 2017, not all calibration coefficients were stated on Aanderaa 4330 calibration sheets. Fixed ~Sept. 2018.

Optode 4330 SN range ~2400 - ~3000

-> AO 103, IF 88, IN 10, HZ 7, JA 1 float

New deployments with 4330 optodes in this SN range need checking!



CALIBRATION CERTIFICATE

Form No 830, Juli 2012

Certificate no: 4330_2903_00137192
Foil batch no: 1517M

Product: 4330
Calibration date: 13.11.2017

Serial no: 2903
Page 1 of 2

| Index | Temperature reference(°C) | [O2] Reference(µM) | Temperature raw data(mV) | Phase reading(°) |
|-------|---------------------------|--------------------|--------------------------|------------------|
| 0 | 29.911 | 0.75 | -203.320 | 60.49 |
| 1 | 20.131 | 0.28 | 112.613 | 61.64 |
| 2 | 10.326 | 0.47 | 438.340 | 62.47 |
| 3 | 0.836 | 1.42 | 737.447 | 63.18 |
| 4 | 0.913 | 22.62 | 735.167 | 60.27 |
| 5 | 0.987 | 44.02 | 732.920 | 57.63 |
| 6 | 1.016 | 65.31 | 732.040 | 55.27 |
| 7 | 1.034 | 106.92 | 731.513 | 51.33 |
| 8 | 1.064 | 149.42 | 730.620 | 48.00 |
| 9 | 1.083 | 213.13 | 730.027 | 43.97 |
| 10 | 1.119 | 317.20 | 728.947 | 39.06 |
| 11 | 1.149 | 421.85 | 728.027 | 35.48 |
| 12 | 1.172 | 527.65 | 727.380 | 32.73 |
| 13 | 10.315 | 15.63 | 438.667 | 59.45 |
| 14 | 10.199 | 34.09 | 442.473 | 56.28 |
| 15 | 10.134 | 50.90 | 444.573 | 53.76 |
| 16 | 10.107 | 83.81 | 445.487 | 49.57 |
| 17 | 10.092 | 117.59 | 445.980 | 46.08 |
| 18 | 10.082 | 168.06 | 446.293 | 41.96 |
| 19 | 10.074 | 250.83 | 446.573 | 37.05 |
| 20 | 10.100 | 334.30 | 445.707 | 33.52 |
| 21 | 10.086 | 418.92 | 446.187 | 30.88 |
| 22 | 20.071 | 12.03 | 114.600 | 58.28 |
| 23 | 19.970 | 26.68 | 117.933 | 54.80 |
| 24 | 19.933 | 40.15 | 119.173 | 52.06 |
| 25 | 19.916 | 66.41 | 119.767 | 47.62 |
| 26 | 19.912 | 93.37 | 119.900 | 44.01 |
| 27 | 19.914 | 133.60 | 119.800 | 39.83 |
| 28 | 19.925 | 199.94 | 119.447 | 34.94 |
| 29 | 19.931 | 267.71 | 119.213 | 31.50 |
| 30 | 19.940 | 336.41 | 118.947 | 28.95 |
| 31 | 29.963 | 9.40 | -204.920 | 57.09 |
| 32 | 29.966 | 21.33 | -205.047 | 53.31 |
| 33 | 29.998 | 32.35 | -206.007 | 50.37 |
| 34 | 30.024 | 53.86 | -206.827 | 45.72 |
| 35 | 30.042 | 75.83 | -207.380 | 42.03 |
| 36 | 30.054 | 108.90 | -207.773 | 37.81 |
| 37 | 30.063 | 163.33 | -208.027 | 33.02 |
| 38 | 30.068 | 219.35 | -208.207 | 29.69 |
| 39 | 30.071 | 277.09 | -208.287 | 27.24 |



CALIBRATION CERTIFICATE



Form No 830, Juli 2012

Certificate no: 4330_2903_00137192
Foil batch no: 1517M

Product: 4330
Calibration date: 13.11.2017

Serial no: 2903
Page 2 of 2

| Index | 0 | 1 | 2 | 3 | 4 | 5 | 6 |
|-------------|-------------|--------------|-------------|--------------|--------------|-------------|------------|
| SVUFoilCoef | 2.80898E-03 | 1.17750E-04 | 2.39215E-06 | 1.25352E02 | -2.27661E-01 | -3.14673E01 | 2.48711E00 |
| TempCoef | 2.35445E01 | -3.05747E-02 | 2.77781E-06 | -4.16717E-09 | 0.00000E00 | 0.00000E00 | |



Date:13.11.2017

Tor-Ove Kvalvaag
Tor-Ove Kvalvaag, Calibration Engineer

Follow-up: PhaseCoef0 issue – post

Following modification of cal. procedures after Apr. 2017, not all calibration coefficients were stated on Aanderaa 4330 calibration sheets. Fixed ~Sept. 2018.

Optode 4330 SN range ~2400 - ~3000

-> AO 103, IF 88, IN 10, HZ 7, JA 1 float

New deployments with 4330 optodes in this SN range need checking!

AANDERAA

a xylem brand

Certificate no: 4330_3023_00151937

Foil batch no: 1517M

Product: 4330

Calibration date: 26.09.2018

Serial no: 3023

Page 1 of 2

| Index | Temperature reference(°C) | [O2] Reference(µM) | Temperature raw data(mV) | Phase reading(°) |
|-------|---------------------------|--------------------|--------------------------|------------------|
| 0 | 30.457 | 2.65 | -205.107 | 58.97 |
| 1 | 20.174 | 2.04 | 124.520 | 60.07 |
| 2 | 10.189 | 2.16 | 453.520 | 60.80 |
| 3 | 0.837 | 2.57 | 745.127 | 61.41 |
| 4 | 0.896 | 20.94 | 743.373 | 58.92 |
| 5 | 0.956 | 43.15 | 741.613 | 56.25 |
| 6 | 1.003 | 64.78 | 740.220 | 53.91 |
| 7 | 1.049 | 107.52 | 738.893 | 49.92 |
| 8 | 1.083 | 151.55 | 737.847 | 46.52 |
| 9 | 1.106 | 216.26 | 737.187 | 42.48 |
| 10 | 1.113 | 319.46 | 736.993 | 37.66 |
| 11 | 1.122 | 427.86 | 736.673 | 33.97 |
| 12 | 1.172 | 532.18 | 735.220 | 31.26 |
| 13 | 10.775 | 16.96 | 434.447 | 57.80 |
| 14 | 10.694 | 34.14 | 437.073 | 54.84 |
| 15 | 10.626 | 50.50 | 439.273 | 52.38 |
| 16 | 10.558 | 83.55 | 441.467 | 48.16 |
| 17 | 10.517 | 117.85 | 442.840 | 44.61 |
| 18 | 10.479 | 169.61 | 444.060 | 40.39 |
| 19 | 10.457 | 254.02 | 444.753 | 35.39 |
| 20 | 10.433 | 335.26 | 445.540 | 31.98 |
| 21 | 10.421 | 420.22 | 445.927 | 29.31 |
| 22 | 20.308 | 13.66 | 120.080 | 56.84 |
| 23 | 20.305 | 27.58 | 120.200 | 53.56 |
| 24 | 20.307 | 41.32 | 120.093 | 50.76 |
| 25 | 20.313 | 67.73 | 119.900 | 46.29 |
| 26 | 20.321 | 94.62 | 119.633 | 42.68 |
| 27 | 20.330 | 135.99 | 119.400 | 38.36 |
| 28 | 20.335 | 202.15 | 119.200 | 33.49 |
| 29 | 20.343 | 269.13 | 118.907 | 30.06 |
| 30 | 20.349 | 338.33 | 118.733 | 27.49 |
| 31 | 30.331 | 11.28 | -201.247 | 55.83 |
| 32 | 30.322 | 22.29 | -200.967 | 52.40 |
| 33 | 30.322 | 34.12 | -200.993 | 49.25 |
| 34 | 30.326 | 55.51 | -201.093 | 44.64 |
| 35 | 30.333 | 78.37 | -201.300 | 40.79 |
| 36 | 30.335 | 112.72 | -201.387 | 36.43 |
| 37 | 30.350 | 166.83 | -201.813 | 31.69 |
| 38 | 30.359 | 222.92 | -202.100 | 28.36 |
| 39 | 30.363 | 281.13 | -202.207 | 25.89 |

AANDERAA

a xylem brand

CALIBRATION CERTIFICATE

Form No 830, Juli 2012

Certificate no: 4330_3023_00151937
Foil batch no: 1517M

Product: 4330
Calibration date: 26.09.2018

Serial no: 3023
Page 2 of 2

Giving these coefficients

| Index | 0 | 1 | 2 | 3 | 4 | 5 | 6 |
|-------------|-------------|--------------|-------------|--------------|--------------|-------------|------------|
| SVUFoilCoef | 2.72541E-03 | 1.20870E-04 | 2.08391E-06 | 1.65032E02 | -1.95740E-01 | -3.55476E01 | 3.26582E00 |
| TempCoef | 2.39762E01 | -3.08300E-02 | 2.86136E-06 | -4.23999E-09 | 0.00000E00 | 0.00000E00 | |

Oxygen validation

x O₂ deviation — +2.5µM/+1.5% — -2.5µM/-1.5%

The graph displays the oxygen concentration in µM on the x-axis (ranging from 0 to 500) and the deviation in µM on the y-axis (ranging from -8 to 8). Blue 'x' marks represent individual O₂ deviation measurements. Two red lines represent the acceptance limits: +2.5µM/+1.5% (upper limit) and -2.5µM/-1.5% (lower limit). All data points fall within the acceptance limits.

| µM (x-axis) | O ₂ deviation (µM) (y-axis) |
|-------------|--|
| 10 | -1.5 |
| 20 | -1.2 |
| 30 | -1.8 |
| 40 | -1.5 |
| 50 | -1.2 |
| 60 | -1.5 |
| 70 | -1.2 |
| 80 | -1.5 |
| 100 | -1.2 |
| 150 | -1.5 |
| 200 | -1.2 |
| 250 | -1.5 |
| 260 | -1.2 |
| 270 | -1.5 |
| 280 | -1.2 |
| 290 | -1.5 |
| 300 | -1.2 |
| 380 | -1.5 |
| 450 | -1.2 |
| 480 | -1.5 |

Temperature validation

x Temp. deviation — +0.03°C — -0.03°C

The graph displays the temperature in °C on the x-axis (ranging from 0 to 30) and the deviation in °C on the y-axis (ranging from -0.04 to 0.04). Blue 'x' marks represent individual temperature deviation measurements. Two red lines represent the acceptance limits: +0.03°C (upper limit) and -0.03°C (lower limit). All data points fall within the acceptance limits.

| °C (x-axis) | Temp. deviation (°C) (y-axis) |
|-------------|-------------------------------|
| 1 | -0.015 |
| 2 | -0.012 |
| 3 | -0.018 |
| 4 | -0.015 |
| 5 | -0.012 |
| 6 | -0.015 |
| 9 | -0.012 |
| 16 | -0.015 |
| 20 | -0.012 |
| 26 | -0.015 |

With following settings

| Index | 0 | 1 | 2 | 3 |
|-----------|--------------|------------|------------|------------|
| PhaseCoef | -4.12000E-01 | 1.00000E00 | 0.00000E00 | 0.00000E00 |

| Index | 0 (Offset) | 1 (Slope) |
|----------|------------|------------|
| ConcCoef | 0.00000E00 | 1.00000E00 |
| Salinity | 0.00 | |


Date:26.09.2018

Tor-Ove Kvalvaag
Tor-Ove Kvalvaag, Calibration Engineer

Optode biofouling

Avoid continuous daily surfacing during the day in warm areas!
Risk of severe biofouling

- daily during noon




Optode biofouling

Avoid continuous daily surfacing during the day in warm areas!

Risk of severe biofouling

- daily during noon
- daily during noon




Optode biofouling

Avoid continuous daily surfacing during the day in warm areas!

Risk of severe biofouling

- daily during noon
- daily during noon
- 4 rapid cycles at 10 d interval




Optode biofouling

Avoid continuous daily surfacing during the day in warm areas!

Risk of severe biofouling

- daily during noon
- daily during noon
- 4 rapid cycles at 10 d interval
- 10 d during noon




Optode biofouling

Avoid continuous daily surfacing during the day in warm areas!

Risk of severe biofouling

- daily during noon
- daily during noon
- 4 rapid cycles at 10 d interval
- 10 d during noon
- 5 d interval



Improve RTQC: Proposal to add Surface Saturation test

Test Definition

- ⇒ Convert DOXY into DOXY saturation
- ⇒ Mean in the first 10 meters (**without levels with QC=4 at range test**)
- ⇒ If $\text{mean}(\text{DOXY_sat}) \geq 150$: QC = 4 for the whole profile
- ⇒ If triggered for 10 consecutive profiles: Put float on greylist with QC = 4

Evaluation

This test was run on the entire Coriolis DOXY fleet on the raw DOXY values (neglecting DOXY_QC)

- ⇒ 29 Coriolis floats trigger the saturation test

Comparison with DM and inputs from greylist yields:

- ⇒ 22 floats were already greylisted by operators with QC = 4
(**with a date prior the trigger => test tends to be conservative**)
- ⇒ 3 floats trigger the test but with DOXY_QC = 4 already set by the range test
(**the saturation test should be performed after the range test**)
- ⇒ 4 floats trigger the test intermittently: Expert feedback confirmed weird sensor behaviour and probably bad, uncorrectable data; or biofouling shortly before float death




DM methods for DOXY adjustment

Sources of reference data:




Direct in air
measurements

- Most accurate reference
- Can remove bias and drift
- No reliance on gridded O₂ fields
- SCOR WG 142 recommendation
- SBE63 optodes not capable of in air measurements
- Requires good DAC decoding (b-traj)
- Artefacts with bad O₂-T-calibration



Surface O₂ saturation
climatology

- Possible with all sensors
- Can remove systematic bias (and drift) but less accurate
- Possible option for RT
- Reliance on match with gridded field



Winkler-calibrated CTD-O₂
deployment profile

- Possible with all sensors
- Can only remove initial bias
- May cover range of O₂ / T
- Spatio-temporal match
- Only one profile; bad statistics



Deep O₂ concentration
climatology / stable layer

- Complement to surface O₂
- Requires stable deep layer
- Reliance on match with gridded field
- High uncertainty (esp. for surface O₂)

Bittig et al. (2018), Johnson et al. (2017), Bushinsky et al. (2015), Johnson et al. (2015), Bittig and Körtzinger (2015), Takeshita et al. (2013), Tengberg et al. (2006)

DM methods for DOXY adjustment

Ways to adjust sensor data:

Optodes

(Aanderaa, SBE63, JAC, ...)

- gain factor on PPOX_DOXY
(+small zero offset if warranted;
order 5 hPa max.)




- Follows sensing principle
- Reliance on O₂-T-calibration
(multipoint !)

Electrodes (SBE43)

- gain factor and offset on DOXY

- Follows sensing principle

No matter where the reference data come from!



DOXY adjustments: Math (Bullet-proof version)

Oxygen optodes sense the O_2 partial pressure *inside* the sensing membrane.

-> Correction must be done *on PPOX_DOXY at inside membrane conditions!*

```
% convert DOXY into DOXY_in_molar_units units conversion (micromol/kg to micromol/L)
rho = potential_density(presValues, tempValues, psalValues);
molarDoxyValues = doxyValues .* rho;

% pressure effect un-correction:
% at presValue, optode quenched by different pO2 inside membrane than pO2 outside in seawater due to re-equilibration effect
% translate already corrected value (outside conditions) back to sensed value (inside membrane)
oxygenPresUncomp = calcoxy_presuncomp(molarDoxyValues, presValues, tempValues, pCoef2, pCoef3);

% convert DOXY_in_molar_units_and_inside_conditions into PPOX_DOXY units conversion (micromol/L to hPa)
ppoxDoxyValues = O2ctoO2p(oxygenPresUncomp, tempValues, psalValues, presValues);

% adjust PPOX_DOXY
ppoxDoxyAdjValues = ppoxDoxyValues * a_slope + a_offset;

% convert PPOX_ADJUSTED into DOXY_ADJUSTED_in_molar_units_and_inside_conditions units conversion (hPa to micromol/L)
oxygenAdjPresUncomp = O2pttoO2c(ppoxDoxyAdjValues, tempValues, psalValues, presValues);

% pressure effect re-correction:
% at presValue, optode quenched by different pO2 inside membrane than pO2 outside in seawater due to re-equilibration effect
% translate adjusted sensed value (inside membrane) to adjusted corrected value (outside conditions)
molarDoxyAdjValues = calcoxy_prescomp(oxygenAdjPresUncomp, presValues, tempValues, pCoef2, pCoef3);

% convert DOXY_ADJUSTED_in_molar_units into DOXY units conversion (micromol/L to micromol/kg)
doxyAdjValues = molarDoxyAdjValues ./ rho;
```

apply a small zero offset only if warranted (batch calibrated optodes?)
order 5 hPa max.

DOXY adjustments: Error estimate


Oxygen optodes sense the O_2 partial pressure *inside* the sensing membrane.

-> Error should be expressed as PPOX_DOXY converted back to DOXY

QC manual recommendation:

Error should be a constant PPOX_DOXY on entire profile, not a fraction of the value

(Work in progress: Complete error budget for DOXY)



```
% convert PPOX_ADJUSTED into DOXY_ADJUSTED_in_molar_units_and_inside_conditions units conversion (hPa to micromol/L)
oxygenAdjPresUncomp = O2ptoO2c(ppoxDoxyAdjValues, tempValues, psalValues, presValues);

% pressure effect re-correction:
% at presValue, optode quenched by different pO2 inside membrane than pO2 outside in seawater due to re-equilibration effect
% translate adjusted sensed value (inside membrane) to adjusted corrected value (outside conditions)
molarDoxyAdjValues = calcoxy_prescomp(oxygenAdjPresUncomp, presValues, tempValues, pCoef2, pCoef3);

% convert DOXY_ADJUSTED_in_molar_units into DOXY units conversion (micromol/L to micromol/kg)
doxyAdjValues = molarDoxyAdjValues ./ rho;
```





DOXY adjustments: Error estimate, 2nd level



- Only one DM method: Following sensing principle
- Accuracy depends on reference data accuracy and character





Direct in air
measurements




Surface O₂ saturation
climatology



Winkler-calibrated CTD-O₂
deployment profile



Deep O₂ concentration
climatology / stable layer



Adjustment workflow

Bittig et al. (2019) A BGC-Argo guide: Planning, deployment, data handling and usage. Front. Mar. Sci. 6:502.
doi: 10.3389/fmars.2019.00502


Float PI

- coordinates with a DAC on upcoming float deployment
- notifies float with AIC
- deploys float




Float

transmits
raw data



DAC

- processes raw data in real-time and
- sends 'R'-files to GDAC within 24 h



GDAC

- performs higher-level aggregation and
- makes files available for users




Users

may incorporate DMQC
results for adjustments
into real-time data stream



DAC

- checks format conformity to Argo rules with file checker
- sends 'D'-files to GDAC to replace real-time 'R' files




Float PI / DMQC expert

- 'Operator analysis':
- assesses float data and
 - provides DMQCD and adjusted data files to DAC (1 or more parameters)



Adjustment workflow



Bittig et al. (2019) A BGC-Argo guide: Planning, deployment, data handling and usage. Front. Mar. Sci. 6:502.
doi: 10.3389/fmars.2019.00502



Time response correction


- Requires expert knowledge / sensor understanding

6901180





Time response correction

- Requires expert knowledge / sensor understanding




Bittig and Körtzinger (2017) Technical note:
Update on response times, in-air measurements, ... Ocean Sci.



Determining Sensor Response-time *in-situ*

- Upon deployment, floats were programmed to profile continuously, recording data including timestamps on both up- and down-casts.
- Optode response time is determined by performing the correction for a range of τ values on pairs of time-adjacent up- and down-casts.
- The value of τ that minimizes the mismatch (Root Mean Square Deviation, RMSD) between the two casts is taken to be the effective sensor response time.



Determining Sensor Response-time *in-situ*

- Median values of effective response times are around 70 s with standard deviations up to 20 s (see Table)
- Large standard deviations in response times likely due to environmental factors rather than changes to the inherent sensor response time – this method determines the effective response time which is highly dependent on the flow at the sensor interface.
- Results are on the low end or just outside the estimated range of response times for autonomous floats provided in Bittig and Körtzinger (2017) of 70-140 seconds, using boundary layer estimates based on float profiling velocity.

| Float ID | τ (s) | σ_{τ} (s) |
|----------|------------|---------------------|
| f7939 | 76.5 | 8.5 |
| f7940 | 69 | 15.5 |
| f7941 | 67 | 16.9 |
| f7942 | 64.5 | 14.9 |
| f7943 | 75 | 16.5 |
| f7944 | 76 | 13.9 |
| f7945 | 73.5 | 13.3 |
| f8081 | 68 | 16.0 |
| f8082 | 73 | 18.8 |
| f8083 | 66.5 | 20.4 |
| Average | 70.9 | 15.5 |

Time response correction

- Requires expert knowledge / sensor understanding
- Currently a DM task


4.6 Further delayed-mode tasks for DOXY

In addition to O₂ adjustment as suggested above and general visual QC, delayed-mode QC for DOXY should try to address:

- Correction of the oxygen optode time response if time stamps are available (see e.g., Bittig and Körtzinger 2017, Bittig et al. 2014).
- If no time stamps are available, correction of the oxygen optode time response may still be feasible using a mean float ascent velocity as input (see figures 2/3 of <https://dx.doi.org/10.5194/os-13-1-2017> and <https://dx.doi.org/10.5194/os-2016-75-AC1>, respectively).
- The presence of a so-called 'O₂ hook' at the base of the profile (first ~50 dbar), i.e., significantly lower O₂ observations as suggested by the O₂ gradient at slightly shallower depths / as indicated from climatological data

Time response prerequisite: Time!


1. Float does not transmit timestamps
Need to get creative (~ascent speed) -> Ken
2. Float transmits timestamps
 - (a) intermittently (-> traj), or
 - (b) at all points (-> traj and profile **"MTIME"**)



saw a need to be able to store timing information in the float's profile file. Following discussions at ADMT16, AST17, and ADMT17 it was concluded that timing information associated with each measured parameter of the profile is scientifically useful for Biogeochemical-Argo, but does not seem necessary for core Argo. Sparse timing data should remain in the trajectory file – it was designed to handle this data and keeps QC of timing information mostly to one file. Abundant timing information (i.e., timing information for all profile observations), however, are stored in the i-parameter "MTIME" in the b-profile files.

Time response prerequisite: Time!

Very few floats with (optional) MTIME in profile



- Presently only filled by Coriolis (41 floats)
- Makes time response correction by an (expert) user very hard!!

Please check your BGC float decoders: Is there timing information that could / should be stored in “MTIME” ?

| | |
|----------------------|---|
| argo data management | Argo data management DOI: http://dx.doi.org/10.13155/47998 |
| | Processing BGC-Argo measurement timing information at the DAC level Version 1.0 March 28 th 2017 |

The most plausible explanation would be an effect of **turbulence on the foil of the optode**

Tests in lab with 3 optodes (from D.Lefevre M.I.O.):

Thermal conditions are stable ($\pm 0.5^{\circ}\text{C}$) and the O_2 forcing condition is maintained at $\pm 0.1 \mu\text{M}$.

The rotation of the blades at 120 rpm is obtained from the start

We can observed:

- The transition from 0 to 120 rpm and 120 to 0 rpm results in a **concentration change of $0.8 \mu\text{M}$** with almost no thermal evolution
- **"Equilibrium" is found after about a hundred minutes**
- There is symmetry in the optode response

Turbulent regime changes near the optode can therefore impact a large part of the DOXY profile

How to correct the bottom DOXY profiles ? How to qualify this part of the profile in a rigorous way (and not only visually) ? How could we reduce the deviation between up & downcasts ?

